



International Quality And Accreditation Services Pvt. Ltd.
(Formerly International Quality And Accreditation Services LLP)
25/33, 2nd Floor, East Patel Nagar New Delhi, 110008

Scope of Accreditation for calibration

As per ISO/IEC 17025:2017

CAB Name Perfect Calibration Centre

Page No. **1** of **1**

Address K73, Udyog Nagar, Rohtak road, Delhi – 110041

Issue date: 05.10.2024

Certificate No C-0007

Validity: 04.10.2028

Amendment date: 16.04.2025
(Re-Issued)

Sr. No.	Parameter/Measurand quantity, Instrument or gauge	Range	CMC	Remarks/Method used	Facility
Fluid Flow					
1.	Fluid Flow/ Quantity by Mass for Flow meter of CNG at pressure range 70-200 bar (Gas medium)	1 kg to 12 kg	+/- 0.75%	Weighing method as per ISO 4185	Permanent

Note: CMC in ($\pm$) is Calibration Measurement Capability expressed as the expanded uncertainty having a coverage factor $k = 2$ at confidence interval to approximately 95%.

This is annexure to 'Certificate of Accreditation' and does not require any signature.

Regd. Office: No. 307/20, 2nd Lane No. 5A, Ranjit Nagar, New Delhi 110008, India