



**International Quality And Accreditation Services Pvt. Ltd.**  
(Formerly International Quality And Accreditation Services LLP)  
25/33, 2nd Floor, East Patel Nagar New Delhi, 110008

### Scope of Accreditation for Calibration

As per ISO/IEC 17025:2017

**CAB Name:** TAYALTECH

Page No. **1** of **1**

**Address:** C-39, 2<sup>nd</sup> Floor, Tayal Technology Hub, Vyapar Vihar,  
Bilaspur- 495001, Chhattisgarh

**Issue date:** 29.04.2025

**Validity** : 28.04.2029

**Certificate No.:** C-0010

**Amendment date:** NA

| Sr. No.               | Parameter/Measurand quantity, Instrument or gauge | Range      | CMC           | Remarks/Method used                     | Facility  |
|-----------------------|---------------------------------------------------|------------|---------------|-----------------------------------------|-----------|
| Calibration- Chemical |                                                   |            |               |                                         |           |
| 1.                    | Breath Analyzer                                   | 80mg/100ml | 6.28 mg/100ml | Wet analysis – Ethanol+ Distilled water | Permanent |

Note: CMC in (±) at 95% confidence level